

In continuation with request Prebid 17/2/28

Changes require in Specifications of MRI Machine

1. Pelvic dedicated should change to Pelvic dedicated/Pelvic or equivalent.
2. Two power generators required.
3. Coil Management system not mandatory (Confusion line).
4. Spine Coil required 32 channels or better.
5. Eight channel coil or coil combination will be acceptable in Small FOV Coil.
6. Slice thickness 200 MM or better will be acceptable.

Homogeneity 3.1 ppm or better

Mammography
↓

Made in France + Italy HD

1 doctor foreign training
for biopsy

↓
Rin

↓
HD