

In continuation to the Pre-bid Minutes 17-02-2018

Corrigendum uploaded

1. France and Nether land will be acceptable as COUNTRY OF ORIGIN in ANGIOGRAPHY.
2. FRANCE will be acceptable in Hematology Analyzer as Country of Origin.
3. In country of Origin for Laundry'' France will be acceptable''.
4. In Operation Theaters OT Table Country of origin' 'Finland'' will also be acceptable.
5. "Turkey will not be considered in CSSD as Country of Origin".
6. CZT based technology will also be acceptable in GAMMA CAMERA.
7. Rotating/Fixed Detectors are also acceptable in GAMMA CAMERA.
8. Simultaneous isotopes scanning will also be acceptable in GAMMA CAMERA.
9. Tungsten Collimator will also be acceptable in GAMMA CAMERA.
10. In country of Origin Switzerland will also be acceptable in GAMMA CAMERA.

