



	or more qty:1	"16 element/ channel or more	agree.
8	Spine Coil: 32 channels for optimum coverage of cervical, thoracic and lumbar spine(embed/ fix not detachable from table) qty:1	Kindly amend to read as "element/ channels And Kindly amend to read as embedded / integrated	Spine coil: 32 channels embedded / integrated both will be acceptable.
9	Dedicated shoulder coil 16 channel qty:1	Kindly amend to read as " Flexible coil with placement pads"	Not agree with the reservation we need specified dedicated coil for shoulder.
10	Breast coil: 16 channels or more with simultaneous imaging of both breasts in all directions. Qty:1	Kindly amend to read as " 8 Elements/ channels"	Not agree with the reservation
11	Latest Air coil	Kindly amend to read as "light weight coil"	Latest AIR coil line removed by the committee for which all bidders agreed anonymously
12	Diffusion tensor Imaging at least 200 direction or more directions	Kindly amend it to read as 190 or more Directions	DTI at least 200 or more directions removed as per discussion and agreed upon by the bidders
13.	Zoom diffusion or Equivalent	Kindly add" Small FOV/Zoom it/ Focus" as equivalent software of all other vendors	Agreed with the reservation the spec already mentioned in the that equivalent software will also be acceptable
14	Product Certification : USFDA, CE, JIS	Along with FDA/CE/JIS also MHLW and PAL should be added	JIS along with MHLW and PAL certificate will also be accepted NOT AGREED
15	Quoted Model performance Certificate : 6 marks 03 performance certificate (preferably local, but international will be acceptable, too) of the quoted model: 02 marks for each certificate	Local and international performance certificate should be given separate marks.	Not agreed by the committee, the SBD Criteria Point will remain the same.
16	Warranty period extension (Beyond 05 Years mandatory warranty)=08 marks 05+01 years=04 marks (02-Marks) 05+02 years= 02 marks (04-Marks) 05+03 years= 04 marks (08-Marks)	The warranty should be defined either comprehensive with coils, helium & cold head replacement or only system warranty.	Extended warranty will be comprehensive warranty and cover all the services provide during the warranty period.
17	Date of submission 31/12/2019	Tender submission date should be extended to 15 th January 2020 as the manufacturer rarely response due to New Year holidays	The point found valid and will be discuss with authorities.
18	Fringe filed (radial x axial) 5.0 x 3 m or better at 0.5 mT	Fringe filed (radial X axial)2.4 x 3.8 or better at 0.5 m T	The reservation found vendor specific, so not entertained and committee

NOT AGREED
by BOG

[Handwritten signature]